

Fillmaster[®]

3600

PHARMACEUTICAL WATER DISPENSER

Quick Reference Guide

To change mixing wait timeout:

Start



Program X5



2

To add a speed key:

Start



Program X6



2

To use a speed key:

Start



0

To check power supply level:

Start



Program X3



2

To recalibrate:

Start



Program X4



1

Support or battery/filter replacements:

Fillmaster Systems, LLC

619-596-9900

www.FillmasterSystems.com